



شركة مصفاة البترول الاردنية المساهمة المحدودة

عمان - الأردن

مناقصة: - (148 / 2025) رقم طلب المواد (25010958) تحديث خلايا التوزين للقبانات ذوات الأرقام 3+2 / دائرة مستودعات العقبة.

الموضوع: - دعوة للاشتراك في مناقصة محلية

ترغب شركة مصفاة البترول الاردنية بإنجاز الاعمال المبينة تفصيلها ومواصفاتها الفنية في الكشف المرفق وباعتباركم متخصصين في هذه الاعمال نوجه هذه الدعوة لكم للاشتراك بالمناقصة اعلاه وتقديم عرض سعر للاعمال المطلوبة إذا وجدتم ان بإمكانكم تنفيذها مع الالتزام بكافة الشروط والمواصفات وفيما يلي الشروط العامة للمناقصة: -

1. يقدم العرض (نسختين) قبل الموعد المحدد ادناه ضمن ظرف مغلق ومختوم مبينا عليه رقم المناقصة وتاريخ الاغلاق واسم وعنوان المناقص ويسلم باليد الى (دائرة المشتريات الإدارة العامة - جبل عمان - الدوار الأول) ولا يقبل اي عرض يرسل بالفاكس نهائيا.

2. يجب ان يكون العرض مروساً باسم وعنوان المورد موقعا ومختوما بالخاتم الرسمي للمناقص ومرفقا بصورة الشروط موقعة ومختومة منه لتأكيد التزامه بكافة الشروط والمواصفات وكذلك تقديم نسخة من السجل التجاري ورخصة المهن.

3. على المناقص تثبيت الاسعار الافرادية والاجمالية على عرضه رقما وكتابة بخط واضح دون تصحيح او كشط او محو.

4. على المناقص بيان الامور التالية في عرضه بوضوح: -

- شمول السعر لضريبة المبيعات 16% او اية نسبة تحدد بعرض السعر والرسوم الجمركية.
- في حال وجود اي خدمة اضافية يراد تقديمها في العرض وغير واردة في مواصفات طلب المواد او امر الشراء يجب ان يتم ذكرها بالتفصيل وان يتم بيان كلفتها المالية بشكل واضح.
- مدة صلاحية الاسعار (يجب الا تقل عن (90) يوم من تاريخ تقديم العرض وخلاف ذلك يحق للمصفاة استبعاد العرض).

- اجور النقل والتركيب الى دائرة مستودعات العقبة يجب ان تكون مشمولة ضمن السعر المعروف.
- ان المواد المعروضة جديدة وغير مستعملة او مجددة وخالية من العيوب مع بيان سنة الصنع.
- ان يحدد منشأ البضاعة والالتزام بتقديم شهادة بلد المنشأ وتحديد اسم الشركة الصانعة، مده التوريد، مكان التسليم علما ان مكان التسليم دائرة مستودعات العقبة.
- تثبيت السعر الافرادي والاجمالي رقما وكتابه بخط واضح دون شطب او تعديل.

5. على المناقص تقديم شهادة منشأ عند توريد المواد (مصدقة اصولا او تشير للبيان الجمركي)

6. الشركة غير ملزمة بالإحالة على اقل الاسعار ويحق لها رفض اي عرض دون ابداء الاسباب.

7. يحق للشركة الغاء المناقصة او اعادة طرحها او شراء جزء منها او التعديل على الكميات المطلوبة دون بيان الاسباب ودون ان تتحمل اي مسؤولية أي كان نوعها.

8. على المناقص عدم مراجعة اي جهة فنية في الشركة الا بالتنسيق مع دائرة المشتريات.

9. المواد المطلوبة خاضعة للزيادة والنقصان بنسبة $\pm 25\%$.

س

2

10. يلتزم المناقص بدفع قيمة طوابع الواردات والبالغة (6) بالألف على قرار الاحالة خلال عشرة ايام من تبليغه بالقرار.

11. على المناقص تقديم كفالة صيانة غير مشروطة بنسبة 5% من القيمة الاجمالية (بعد التشغيل الناجح) ولمدة ثلاث سنوات شاملة قطع الغيار والايدي العاملة.

12. يتم دفع ثمن الاعمال المطلوبة بعد انجاز الاعمال والموافقة عليها من قبل دائرة مستودعات العقبة وبعد صدور سند استلام يؤكد استلامها كاملة ومطابقة للشروط والمواصفات.

13. في حال تأخير توريد / تنفيذ المناقصة حسب ما تم الاتفاق عليه يترتب عليكم غرامة قدرها (20) دينار أردني عن كل يوم تأخير.

14. في حال اخلال المتعهد المحال عليه المناقصة باي من التزاماته، يحق للشركة الغاء الاحالة واللجوء الى طرف ثالث دون الحاجة لأي اخطار او انذار عادي او عدلي ودون الحاجة لمراجعة القضاء وتحميل المتعهد كافة التكاليف المترتبة على ذلك

15. تعتبر هذه الشروط جزءاً لا يتجزأ من المناقصة وعلى المورد الالتزام بما جاء فيها.

16. اخر موعد لقبول واستلام العروض الساعة الثالثة من بعد ظهر يوم الخميس الموافق 2025/11/13 وتسلم لدائرة

المشتريات - الادارة العامة - جبل عمان - الدوار الأول - الطابق الأول - مكتب 10

17. في حال وجود أي استفسار يتم ارساله بالبريد الإلكتروني على العنوان (Local.Purchasing@jopetrol.com.jo).

18. يجب ان تكون الفاتورة المقدمة من قبلكم في حال شراء السلع او تقديم خدمة وفقاً لنظام تنظيم شؤون الفوترة والرقابة عليها رقم (34) لسنة 2019 وسيتم رفض أي فاتورة لا تكون وفقاً للنظام أعلاه.

كما يشترط تقديم البيانات التالية على الفواتير الصادرة من قبلكم: -

الاسم الكامل للشخص أو الشركة أو المفوض بالتوقيع وختم الشركة الرسمي أو المفوض بالتوقيع وتوقيعه.

الرقم الوطني الكامل للشخص أو الشركة أو المفوض بالتوقيع.

العنوان الكامل بالإضافة إلى رقم الهاتف والصندوق البريدي والرمز البريدي.

الرقم الضريبي للشخص أو الشركة إن وجد.

504

504

Jordan Petroleum Refinery Company
Quote Order

Order Number 25010498 OQ 00001
Branch/Plant F612706

Order Date 30/09/2025 Request Date 30/09/2025
Freight Handling

Delivery Instructions

Line	Item Number	Description	Ordered Quantity	Unit	Supplier Offer Unit Price	Supplier Offer Currency Code	Original Order No	Type	Line No	Orig
1		Upgrading analogue two-scale w	2.000	EA			25010958	OR	1	

Upgrading analogue two-scale weighbridges (No. 2 & 3 at Aqaba Stores Department) to digital weighing system by supplying, installing, commissioning, and calibration by JSMO.
Calibration by Jordan Standards & Metrology Organization (JSMO) shall be included and under responsibility of the contactor.
No mechanical changes are permitted to the scale plates. Any necessary minor adjustments for installation must be formally approved by JPRC.

Each weighbridge (No. 2 and No. 3) contains eight (8) load cells.
Existing load cells (Type RC3-40T C3) shall be replaced with Type RC3d-40T certified digital load cells, maintaining original capacity, tolerance, and calibration accuracy.
Existing Western Scale indicators (S-DF2000) shall be replaced with FLINTEC FT112D Digital Weighing Indicators, ensuring full compatibility with the new load cells and integration with JPRC's ERP system.
The Contractor shall be solely responsible for obtaining all relevant information regarding the existing weighing scale, including its operational integrity and type of connection with JPRC's ERP system.

The Contractor shall formally report all observations, findings, or deviations to JPRC in writing.
The Contractor shall ensure that, upon completion, the new digital weighing scale system is fully operational and integrated, providing the same functionality and continuity of operations as the existing system.
Installation and calibration (by JSMO) for both weighbridges to be completed within max. 5 working days, commencing from the date of site handover to the Contractor.

One-year manufacturer warranty.

Any malfunction within three (3) months from commissioning requiring recalibration of the scale by JSMO shall be addressed by the supplier.

Excludes malfunctions caused by improper use or mechanical shock to the scale.

Supplier shall respond to any JPRC-reported malfunction within 72 hours

Contractor shall supply the following spare parts:

- Load Cell – 4 pcs
- Indicator – 1 pc
- Cable & cable lugs – 4 sets
- End Resistor – 4 units

Quote Order

Order Number 25010498 00001

Line	Item Number	Description	Ordered Quantity	Supplier Offer Unit Price	Supplier Offer Currency Code	Original Order No	Orig TypeLine No
------	-------------	-------------	------------------	---------------------------	------------------------------	-------------------	------------------

الاسعار شاملة التوصيل الى دائرة مستودعات العقبة
الاسعار شاملة التركيب داخل مستودعات العقبة
الاسعار شاملة الضرائب و الرسوم الجمركية و اية رسوم اخرى
الاسعار شاملة الكفالة

Purchasing Manager :

Remarks of Purchasing Department :

.....

.....

Section	Requirement	Details	Compliance	Remarks
Technical Compliance	Objective	Upgrading analogue two-scale weighbridges (No. 2 & 3 at Aqaba-Stores Department) to digital weighing system by supplying, installing, commissioning, and calibration by JSMO.		
	Meteorological / Calibration	Calibration by Jordan Standards & Metrology Organization (JSMO) shall be included and under responsibility of the contractor.		
	Mechanical change	No mechanical changes are permitted to the scale plates. Any necessary minor adjustments for installation must be formally approved by JPRC.		
	Load Cells	- Each weighbridge (No. 2 and No. 3) contains eight (8) load cells. - Existing load cells (Type RC3-40T C3) shall be replaced with Type RC3d-40T certified digital load cells, maintaining original capacity, tolerance, and calibration accuracy.		
	Indicators	Existing Western Scale indicators (S-DF2000) shall be replaced with FLINTEC FT112D Digital Weighing Indicators, ensuring full compatibility with the new load cells and integration with JPRC's ERP system.		
	Gathering Information	- The Contractor shall be solely responsible for obtaining all relevant information regarding the existing weighing scale, including its operational integrity and type of connection with JPRC's ERP system. - The Contractor shall formally report all observations, findings, or deviations to JPRC in writing. - The Contractor shall ensure that, upon completion, the new digital weighing scale system is fully operational and integrated, providing the same functionality and continuity of operations as the existing system.		
	Execution Duration	Installation and calibration (by JSMO) for both weighbridges to be completed within max. 5 working days, commencing from the date of site handover to the Contractor.		
	Warranty	- One-year manufacturer warranty. - Any malfunction within three (3) months from commissioning requiring recalibration of the scale by JSMO shall be addressed by the supplier. - Excludes malfunctions caused by improper use or mechanical shock to the scale. - Supplier shall respond to any JPRC-reported malfunction within 72 hours.		
	Spare Parts Supply	Contractor shall supply the following spare parts: • Load Cell – 4 pcs • Indicator – 1 pc • Cable & cable lugs – 4 sets • End Resistor – 4 units		
Commercial Compliance	Payment Terms	100% after full job completion.		
	Company Details	Company name, full address, offer number.		
	Offer Validity	At least 60 days from closing date.		
	Legal Requirement	In accordance with Jordan prevailing law, supplier shall pay 6 per thousand of the total order value to cover the revenue stamps.		
	Optional / Additional Services	Any extra optional and/or additional services must be clearly detailed and priced.		
	Delivery	Promised service / delivery date to be provided.		
	Price	To be indicated inclusive/exclusive of income tax.		